



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/003,011	11/01/2001	Roy K. Greenberg	PA-5270-RFB	3255
<div>7590 06/27/2007</div> <div>Brinks Hofer Gilson & Lione</div> <div>P.O. Box 10395</div> <div>Chicago, IL 60610</div>				
<div>EXAMINER</div> <div>PHILOGENE, PEDRO</div>				
<div>ART UNIT</div> <div>3733</div>				
<div>PAPER NUMBER</div>				
<div>MAIL DATE</div> <div>06/27/2007</div>				
<div>DELIVERY MODE</div> <div>PAPER</div>				

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Office Action Summary

Application No.

10/003,011

Applicant(s)

GREENBERG ET AL.

Examiner

Pedro Philogene

Art Unit

3733

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 10 April 2007.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1,3-20,22 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☐ Claim(s) _____ is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
2. ☐ Certified copies of the priority documents have been received in Application No. _____.
3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).
- * See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- 1) ☒ Notice of References Cited (PTO-892)
- 2) ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
- 3) ☐ Information Disclosure Statement(s) (PTO/SB/08)
Paper No(s)/Mail Date _____.
- 4) ☐ Interview Summary (PTO-413)
Paper No(s)/Mail Date. _____.
- 5) ☐ Notice of Informal Patent Application
- 6) ☐ Other: _____.

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Claims 1, 4-7 are rejected under 35 U.S.C. 103(a) as being unpatentable over Boyle et al. (6,695,813) in view of Cathcart et al. (5,681,347) in view of Braunschweiler et al. (5,484,444).

With respect to claim 1, Boyle et al., disclose a medical grasping device comprising: an elongate control member (18,520) having an atraumatic distal tip section, as best seen in FIG.1, and a proximal end portion; the elongate control member further including a grasping portion (14,16,530) proximal the distal tip section; an outer sheath (46,48) with a passageway therethrough, as best seen in FIG.2, surrounding the elongate control member and relatively movable with respect thereto.

Although Boyle et al teach of a control assembly, as set forth in column 24, lines 19-45, it is noted that Boyle et al., did not teach of a control assembly as claimed by applicant. However, in a similar art, Cathcart et al., evidences such a control assembly to enable the control deployment and displacement of a device.

Therefore, given the teaching of Cathcart et al., it would have been obvious to one having ordinary skill in the art at the time the invention was made to use the control assembly, as taught by Cathcart et al, in the device of Boyle et al., to urge the medical grasping device from a retracted to an expanded position.

Art Unit: 3733

Furthermore, it is noted that the above combination of references did not teach of an elongated control member that is a flexible cannula defining a lumen extending threthrough into which a guide wire is receivable and movable with respect thereto; as claimed by applicant. However, in a similar art, Braunschweiler et al evidence the use of such an elongated member with cannula and guide wire to ensure that reliable operation is achieved and therefore guaranteed the greatest possible operational reliability.

Therefore, it would have been obvious to one having ordinary skill in the art at the time the invention was made to modify the device of Boyle/cathcart et al, as taught by Braunschweiler et al., to ensure that reliable operation is achieved and therefore guaranteed the greatest possible operational reliability.

With respect to claims 4-7, the above combination of references teaches all the limitations, the outer sheath being flexible and kink-resistant, as set forth in column 11, lines 42-67, column 12, lines 1-33, the atraumatic distal tip section tapers to a blunt and rounded tip; as best seen at the end of the control member 18, the control assembly including an actuation section that is grippable for reciprocal movement along the handle, as set forth in column 6, lines 3-25 of Cathcart et al., and a connecting block (25) as set forth in column 6, lines 3-25 of Cathcart et al.

Claim 3 is rejected under 35 U.S.C. 103(a) as being unpatentable over Boyle et al. (6,695,813) in view of Cathcart et al. (5,681,347) in view of Braunschweiler et al. (5,484,444) in view of Gunther et al. (5,330,484).

With respect to claim 3, it noted that the above combination of references did not teach of a hemostatic seal between the sheath and the elongate control member; as claimed by applicant. However, in a similar art, Gunther et al evidence the use of a hemostatic seal to hold the legs of a grid body.

Therefore, given the teaching of Gunther et al., it would have been obvious to one having ordinary skill in the art at the time the invention was made to modify the device of Boyle/Cathcart/Braunschweiler et al., as taught by Gunther et al to provide a hemostatic seal between the sheath and the elongate control member to hold the legs of the grasping portion.

Claims 8-22 are rejected under 35 U.S.C. 103(a) as being unpatentable over Boyle et al. (6,695,813) in view of Cathcart et al. (5,681,347) in view of Hillstead (5,098,440) in view of Braunschweiler et al. (5,484,444).

With respect to claims 8-21, it is noted that the above combination of references teaches all the limitations, except for wire loops that are substantially circular upon full deployment, as claimed by applicant. However, in a similar art, Hillstead evidences the use of wire loops that are circular upon full deployment and having side sections that overlap and touch the vessel wall to engage the object to be retrieved with a greater force.

Therefore, given the teaching of Hillstead, it would have been obvious to one having ordinary skill in the art, at the time the invention was made to incorporate the design of the grasping device of Hillstead in the grasping device of Boyle/Cathcart et al. to engage the object to be retrieved with a greater force.

Furthermore, it is noted that the above combination of references did not teach of an elongated control member that is a flexible cannula defining a lumen extending through into which a guide wire is receivable and movable with respect thereto; as claimed by applicant. However, in a similar art, Braunschweiler et al evidence the use of such an elongated member with cannula and guide wire to ensure that reliable operation is achieved and therefore guaranteed the greatest possible operational reliability.

Therefore, it would have been obvious to one having ordinary skill in the art at the time the invention was made to modify the device of Boyle/cathcart et al, as taught by Braunschweiler et al., to ensure that reliable operation is achieved and therefore guaranteed the greatest possible operational reliability.

Response to Amendment

Applicant's arguments filed 4/10/07 have been fully considered but they are not persuasive. Applicant stated that the cannula of Braunschweiler does not at all relate to the devices of Cathcart or Boyle. The cannula is not attached to either a control assembly or a grasping device. First, applicant's attention is directed to Boyle where an elongated control member, an outer sheath and a control assembly are disclosed. The only difference is that the elongated control member and the control assembly are different as claimed by applicant. However, in a similar art, Cathcart and Braunschweiler, both disclosed a control assembly and an elongated control member, respectively, as claimed by applicant. Therefore, combining the references would have been obvious to one having ordinary skill in the art. Applicant also stated, in addition, no

Art Unit: 3733

factual support is provided in any of the reference that provide a suggestion to combine or indicate that combination would have been obvious to one of ordinary skill in the art at the time of the invention. In response to applicant's argument that there is no suggestion to combine the references, the examiner recognizes that obviousness can only be established by combining or modifying the teachings of the prior art to produce the claimed invention where there is some teaching, suggestion, or motivation to do so found either in the references themselves or in the knowledge generally available to one of ordinary skill in the art. See *In re Fine*, 837 F.2d 1071, 5 USPQ2d 1596 (Fed. Cir. 1988) and *In re Jones*, 958 F.2d 347, 21 USPQ2d 1941 (Fed. Cir. 1992). In this case, the question is not whether the combination was obvious to the applicant but whether the combination was obvious to a person with ordinary skill in the art. Given the problem to be solved, under the correct analysis, any need or problem known in the field and addressed by the patents can provide a reason for combining the elements in the manner claimed. It is common sense that familiar items may have obvious uses beyond their primary purposes, and a person of ordinary skill often will be able to fit the teachings of multiple patents together like a piece of a puzzle. Regardless of Braunschweiler primary purpose, it provide an obvious example of an elongated member that is a flexible cannula defining a lumen extending therethrough into which a guide wire is receivable and movable with respect thereto, and the prior art was replete with patents indicating that such an elongated member was ideal for a medical device. In addition, where there is a design need or market pressure to solve a problem and there are a finite number of identified, predictable solutions, a person of ordinary skill in

Art Unit: 3733

the art has good reason to pursue the known options within his or her technical grasp. The proper question was whether a medical grasping device designer of ordinary skill in the art, facing the wide range of needs created by the developments in the fields, would have seen an obvious benefit to upgrading Boyle with an elongated control member that is a flexible cannula. Braunschweiler taught of an elongated control member that is a flexible cannula to ensure that reliable operation is achieved. The designer, accordingly, would follow the teaching of Braunschweiler. Applicant has not shown anything in the prior art that taught away from the use of Braunschweiler, not any secondary factors to dislodge the determination that at least claim 1 is obvious. Finally, in *Sakraida v. AG Pro, Inc.*, 425 U.S. 237 (1976) the Court derived from the precedents the conclusion that when a patent "simply arranges old elements with each performing the same function it had been known to perform" and yields no more than one would expect from such arrangement, the combination is obvious.

Conclusion

THIS ACTION IS MADE FINAL. Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of

Art Unit: 3733


the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the mailing date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Pedro Philogene whose telephone number is (571) 272-4716. The examiner can normally be reached on Monday to Friday 6:30 AM to 4:00 PM.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Eduardo Robert can be reached on (571) 272 - 4719. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

Pedro Philogene
June 12, 2007


PEDRO PHILOGENE
PRIMARY EXAMINER